

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Barnegat Township Police Department 900 West Bay Avenue Barnegat, NJ 08005		2. Contact: Sgt. Timothy Bradshaw 3. Phone Number: 609-698-5000	4. Date: 05/21/2026 5. County: Ocean
6. Alcotest Instrument Serial Number: ARMK-0008	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service:

The Alcotest instrument was returned from outside evaluation and placed back in service.

10. Comments:

See Draeger Return and Repair form.

Last Known Sequential File #: N/A

- ☐ 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- ☐ 12. The above Instrument/Component is placed out of service pending further evaluation.
- ☒ 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. II David Bellay #8112

Name & Badge Number (Print)


Signature

05/21/2026

Date

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: TOWNSHIP OF BARNEGAT

Date Received: 04/03/2026

Date given to service: 05/14/2026

Carrier: ☐ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☐ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☐ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☒ 9510
☐ 7410 Whole

Serial Number: AR MK 0008

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☒ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414160	A9510 Calibration	1 EA	
4414166	Labor	.5 HR	

Repair Notes: Able to recreate Interferent Detected error. IR Period Avg. adjusted. Internal Standard adjusted.

EC SENSOR =

CAL W/QC AND OPS CHECK

Service Technician LM

Date: 05/14/2026

Delivery receipt



Customer no.
150060023

Number of report
312046181

Please reference on inquiries

Date of receiving Delivery receipt date
02.04.2026 14.05.2026

Consignee, 150060023

**TOWNSHIP OF BARNEGAT
POLICE DEPARTMENT
900 W BAY AVE
BARNEGAT NJ 08005-1298
USA**

Customer, 150060023

TOWNSHIP OF BARNEGAT
900 W BAY AVE
BARNEGAT NJ 08005-1298
USA

Your order

Date of order:
18.03.2026
Purch.ord.no.:
CONTRACT

Branch text

Draeger contact person

CONTRACT OPERATIONS
1-800-437-2437
Contract.operations@draeger.com

Page 1

Quantity	Description	Part no.	Equipment - No.	Serial - No.
000	Alcotest 9510 New Jersey	8327998	1028504910	ARMK-0008
	REPAIR ALCOTEST 9510			

Draeger Inc.
Our Tax ID: 23-1699096
3135 Quarry Road; Telford, PA 18969
An Equal Opportunity Employer M / F / V / H
Telephone 800-437-2437
<http://www.draeger.com>

Remit to:
LOCKBOX (Standard USPS)
Draeger, Inc.
PO Box 13369
Newark, New Jersey
07101-3362

Remit to:
LOCKBOX (Overnight)
FIS - Lockbox Operations
Attn: Draeger, Inc., Lockbox #13369
101 Woodcrest Road
Suite 201
Cherry Hill, New Jersey 08003

Remit US Wire Transfers to:
Account Name: Draeger Inc.
Account Number: 00-494-936
Transit Routing: 021001033
SWIFT: BKTRUS33
Deutsche Bank Trust Company Americas
60 Wall Street 25th Fl, New York, NY 10005

Service note

Customer no. 150060023 Number of report 312046181 Date of receiving 02.04.2026

Please reference on inquiries

Customer
TOWNSHIP OF BARNEGAT
900 W BAY AVE
BARNEGAT NJ 08005-1298
USA

Consignee
TOWNSHIP OF BARNEGAT
900 W BAY AVE
BARNEGAT NJ 08005-1298
USA

Your order

Date of order:
18.03.2026
Your reference:
CONTRACT

NL-Text

Your contact person

CONTRACT OPERATIONS
1-800-437-2437
Contract.operations@draeger.com

Page 1

Item	Quantity	Part no. Description
		Service Order: 155462176
		Able to recreate Interferent Detected er
		Eq.No. : 1028504910 SN: ARMK-0008
		Material: 8327998 Alcotest 9510 New Jersey
		Inv.No. :
		Location:
		Able to recreate Interferent Detected error. IR Period Avg. adjusted.
		Internal Standard adjusted. Calibration complete.
		REPAIR ALCOTEST 9510 10 YR EXT WARR
	0,0 H	

Service Report

Customer No.
0150060023

Contract
317023645

Dispatch No.:
U08340300

Date
05/14/2026

Responsible Service Engineer
Lyfon Morris (5489)

Postal Address

TOWNSHIP OF BARNEGAT

Page 1 / 1

900 W BAY AVE
BARNEGAT, NJ08005-1298
NicholasMimikos

Sold To Party

0150060023
TOWNSHIP OF BARNEGAT

900 W BAY AVE
BARNEGAT, NJ08005-1298

Your order dated

04/02/2026
CONTRACT

Your contact person

Nicholas Mimikos
609-571-2855
609-571-2855

Pos.	Quant.	Part No.	Description	EQ. No. / UDI
		8327998	Alcotest 9510 New Jersey Serial No.: ARMK-0008 S/O 155462176 Return Authorization: 312046181 Inv.No.: Location: Work performed: Able to recreate Interferent Detected error. IR Period Avg. adjusted. Internal Standard adjusted. Calibration complete.	1028504910 /
010	0.50 Hour	R004	Repair Time - Workshop	

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Barnegat Township Police Department 900 West Bay Ave Barnegat, NJ 08005		2. Contact: Sgt. Timothy Bradshaw 3. Phone Number: 609-698-5000		4. Date: 03/16/2026 5. County: Ocean	
6. Alcotest Instrument Serial Number: ARMK-0008		7. Simulator Component Serial Number: N/A		8. Temperature Probe Component Serial Number: N/A	
9. Reason for Service: I arrived at Barnegat Township Police Department to take the Alcotest 9510 (Serial Number ARMK-0008) out of service.					
10. Comments: On March 16, 2026 I arrived at the Barnegat Township Police Department to take the Alcotest 9510 (ARMK-0008) out of service. Operational condition of the instrument was assessed by reviewing the history of multiple Interference error messages (last Sequential file #0269). The instrument will remain out of service at this time and returned to Draeger Inc. for evaluation. Last Known Sequential File #: 0269					
☐ 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.					
☒ 12. The above Instrument/Component is placed out of service pending further evaluation.					
☐ 13. The above Instrument/Component is placed back in service.					
"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."					
Tpr. James Fillimon #8306 Name & Badge Number (Print)		Tpr. James Fillimon #8306 3/16/2026 Signature Date			